

Republic of the Philippines
Department of Science and Technology
PHILIPPINE TEXTILE RESEARCH INSTITUTE
General Santos Avenue, Bicutan, Taguig City
88372071 to 82 local 2360 or 8837-1325
ptri@.ptri.dost.gov.ph
ptridost47@gmail.com

NOTICE OF AWARD

28 December 2021


MR. RECELITO L. POBLETE
President and General Manager
GRYKE SCIENTIFIC INSTRUMENTS TRADING
237-B Worth Drive Street, SAVVY 25 Subdivision,
Severina KM 18, Parañaque City

Dear Mr. Poblete:

We are pleased to notify you that your Bid on December 09, 2021 for execution of the **Procurement of One (1) Lot Respirator Resistance Tester** for the Contract Price of equivalent to **TWO MILLION FIVE HUNDRED EIGHTY SIX THOUSAND TWO HUNDRED EIGHTY EIGHT PESOS (Php2,586,288.00)**, as corrected and modified in accordance with the Instructions to Bidders is hereby accepted.

You are hereby required to provide within ten (10) days the performance security in the form and the amount stipulated in the Instructions to Bidders. Failure to provide the performance security shall constitute sufficient ground for cancellation of the award and forfeiture of the bid security.

Yours sincerely,



CELIA B. ELUMBA
Director IV

Conforme:



MR. RECELITO L. POBLETE
Name of Bidder

Date: 12/28/21



Republic of the Philippines
Department of Science and Technology
PHILIPPINE TEXTILE RESEARCH INSTITUTE
General Santos Avenue, Bicutan, Taguig City, Metro Manila
Tel Nos (632)837-2071 to 82 loc. 2360 to 2379 Telefax (632)837-1325
<http://www.ptri.dost.gov.ph> email: ptri@ptri.dost.gov.ph | ptridost47@gmail.com



BAC Resolution Declaring SCRB and Recommending Approval

RESOLUTION No. PB55- 2021

WHEREAS, the **PHILIPPINE TEXTILE RESEARCH INSTITUTE (PTRI)** advertised the Invitation to Bid for the Procurement of **One (1) Lot Respirator Resistance Tester** charged to CY 2021 GAA-Contingent Fund "Establishment of the DOST-PTRI Medical Textile Testing Center", in the PhilGEPS and DOST-PTRI websites, and posted in a conspicuous place at the premises of PTRI, continuously for seven (7) days;

WHEREAS, in response to the said advertisement and posting, two (2) bidders submitted an Expression of Interest and purchased the bid documents;

WHEREAS, two (2) bids were received before the bid opening on 09 December 2021, 1:00 PM to 2:00 PM and passed the preliminary examination of bid;

Name of Bidder	Bid Amount (As Read)	% Variance from ABC
KRYPTON INTERNATIONAL RESOURCES SALES AND SERVICES, INC.	Php2,318,405.00	14.13%
GRYKE SCIENTIFIC INSTRUMENTS TRADING	Php2,586,288.00	4.21%

WHEREAS, upon evaluation of the technical specifications offered by Krypton International Resources Sales and Services, Inc., it was found to be non-complying on the following technical requirements: I. Mask & Respirator Breathing Resistance Tester - Complies with EN 149 & NIOSH 42 CFR Part 84; Inhalation resistance test flow rate: 30L/min, 85L/min and 95L/min; Exhalation resistance test flow rate: 85L/min and 160L/min; Range of Flow Meter: 0-400L/min, accuracy: 1% or better; Pump capacity: ≥ 300 L/min; f. Power Supply: AC 220V, 60Hz (if not the same, supplier should provide converter); II. Rain tester for Water Resistance Test Nozzle: 13 x \varnothing 0.99mm; III. Washer and Dryer (2 units) Dryer type: Spin Dry; Inverter turbo wash, Steam Technology, Al Tempered Glass Color; Full Stainless-Steel Lifter; Power Supply: AC 220V, 60HZ (if not the same, supplier should provide converter); Includes Inlet hose, User Manual, Spanner, Hose Retainer, Installation; IV. Sewing Machine (2 units) Heavy duty metal frame and stainless-steel bed plate; Includes General purpose sewing foot, pack of needles, bobbins, spool caps, auxiliary, spool pin, spool pin felt. Screwdriver, Seam ripper/Lint brush and soft dust cover; Power Supply: AC 220V-240V, 60 Hz (if not the same, supplier should provide transformer or converter) V. Vacuum cleaner (3 units) Cordless, Convertible handheld and stick vacuum (as shown in picture); Suction power: 30 W or better; Power Supply: Rechargeable battery; Includes battery and charger, nozzles, pipe, filter; VI. Uninterruptible Power Supply (3 units) free slots: 1 or better; Lead acid battery; Other Requirements: Equipment must be part of the regular fleet of the manufacturer and must have already been sold by the manufacturer within the past 2 years with no modifications except for the power supply;

WHEREAS, upon the evaluation of the PTRI Bids and Awards Committee (BAC), a Notice of Post Disqualification was issued on 17 December 2021. This was not acknowledged by Krypton International Resources Sales & Services Inc.

WHEREAS, upon evaluation of the technical specifications offered by Gryke Scientific Instruments Trading its satisfactorily comply with all of the stipulated requirements for the technical specifications.

WHEREAS, the detailed evaluation of bid conducted resulted in the following:

Name of Bidder	Bid Amount (As Calculated)	%Variance from ABC
KRYPTON INTERNATIONAL RESOURCES SALES AND SERVICES INC.	Php2,318,405.00	14.13%
GRYKE SCIENTIFIC INSTRUMENTS TRADING	Php2,586,288.00	4.21%

WHEREAS, upon careful examination, validation and verification of all the eligibility, technical and financial requirements the bid submitted by **GRYKE SCIENTIFIC INSTRUMENTS TRADING**, has been found to be responsive;


NOW, THEREFORE, We, the Members of the Bids and Awards Committee, hereby **RESOLVE** as it is hereby **RESOLVED**:

- a) To declare **GRYKE SCIENTIFIC INSTRUMENTS TRADING**, as the Bidder with a Single Calculated and Responsive Bid (SCRB) for the procurement of **One (1) Lot Respirator Resistance Tester**;
- b) To recommend for approval by the PTRI Director the foregoing findings.

RESOLVED, at the PTRI Training Room, this 27th of December 2021


DONNA A. ULDO
End-user


MERLITA R. ODI
BAC Member


JENNELI E. CAYA
BAC Member



CORAZON I. TAPULGO
BAC Vice-Chairperson


DANIEL LEON J. LAVIN
BAC Member


MARNIE B. DONES
BAC Member


JULIUS L. LEANO, JR., Ph.D.
BAC Chairperson

Approved:


CELIA B. ELUMBA
Director IV

Approved on: 27 December 2021